



INSTITUTE OF NON-FERROUS METALS

Analytical Chemistry Department
44-101 Gliwice, ul. Sowińskiego 5

CERTIFICATE OF ANALYSIS

Chemical reference material of Ag 10/1

<i>Mean oxygen content</i>	<i>11 ppm</i>
<i>Average deviation</i>	<i>3 ppm</i>
<i>Average interlaboratory deviation</i>	<i>5 ppm</i>
<i>Half confidence interval of arithmetic mean</i>	<i>3 ppm</i>

The chemical analyses have been carried out in six laboratories by using of the method of reduction melting in vacuum or in nitrogen atmosphere.

Each package contains two silver rods 6 mm in diameter and 100 mm long.

Director of the Institute

Prof. Ph.D. Zbigniew Śmieszek