

CERTIFICATE OF ANALYSIS

Copper CCA series

The assigned values¹ and uncertainties² in % w/w

Element No.	CCA1		CCA2		CCA3		CCA4		CCA5	
Ag	0.00720	±0.00011	0.0539	±0.0007	0.757	±0.022	1.524	±0.012	1.964	±0.185
Cu	the rest		the rest		the rest		the rest		the rest	

¹ Unweighted mean value of the means of accepted sets of data, each set being obtained in a different laboratory and/or with a different method of determination.

Prof. Zbigniew Śmieszek Director of the Institute

Shows

Certified on October 2018

· Pb

² The certified uncertainty is the expanded uncertainty with a coverage factor k=2, corresponding to a level of confidence of about 95 %.



Description of the material:

The certified reference materials are available in the form of discs (40 mm diameter and ~25 mm height).

Analytical methods applied:

Ag – Inductively coupled plasma optical emission spectrometry (ICP OES), Flame atomic absorption spectrometry (FAAS)

Participants:

Institute of Non-Ferrous Metals, Analytical Chemistry Department, Gliwice, Poland

- Emission Spectrometry Laboratory
- Atomic Spectrometry Laboratory

Walcowania Metali Nieżelaznych "Łabędy" S.A., Gliwice, Poland

Intended use:

The CRM is intended for establishing or checking the calibration of optical emission and X-ray spectrometers for analysis of samples of similar matrix composition (for micro-analysis is not verified).

Instructions for use:

Before every use, the surface of CRM must be prepared by milling or turning on a lathe. Samples should be prepared in the same way as the CRM.

Brief description of the production and certification process:

The CRM_s – CCA were made by melting of all components in the inductive, of crucible furnace and by casting into special moulds protecting elimination of segregation of the components during solidification. Homogeneity testing were made taking into account over 50% of the material produced. Investigations were carried out using atomic emission spectrometry method with low voltage spark. Homogeneity was estimated statistically with application of the test F.

The set consists of 5 certified reference materials in form of discs 40 mm in diameter and ~25 mm height.

The certification of CCA series is valid indefinitely, within the measurement uncertainties specified, provided the CRM is handled in accordance with the instructions given in this certificate.

e Ph