

CERTIFICATE OF REFERENCE MATERIAL



CCC2 Copper alloy

The assigned certified values¹ and uncertainties²

	Pb	Cd	Cr
	%	mg/kg	%
CCC2	0.0987	99	0.0696
	±0.0064	±11	±0.0059

¹ Unweighted mean value of the means of accepted sets of data, each set being obtained in a different laboratory and/or with a different method of determination.

Not certified values

	Hg
	%
CCC2	0.0121 - 0.0382

The information value for Hg is presented as a range because of less homogenic distribution of the element in the material

The certified reference material was developed within the REFROHS project funded by targeted subsidy of Łukasiewicz Research Network and realized by consortium of:

Łukasiewicz Research Network – Institute of Non-Ferrous Metals,

Łukasiewicz Research Network – Institute of Ferrous Metallurgy

Łukasiewicz Research Network – Cracow Institute of Technology

SERTIFIADAMONAVERNOSIE 2002 INSTYTUT METALI NIEŻELAZNY III. Dyrektor Centrum Metalyrgii

dr inz...Tadeusz Gorewoda, MBA Description of the material: Signature

SIEĆ BADAWCZA ŁUKASIEWICZ INSTYTUT METALI WEŻELAZNYCH

DYREKTOR

dr inż. Gartjara Juszczyk

Version 1 from 30th of November 2022

² The certified uncertainty is the expanded uncertainty with a coverage factor k=2, corresponding to a level of confidence of about 95 %.

Description of the material:

The certified reference material is available in the form of discs (40 mm in diameter and ~ 25 mm in height).

Traceability:

The certified values are traceable to the SI via calibration using pure metals, certified monoelement standard solutions or certified reference materials i.e. MBH 39X17869AF, MBH 39X17869AG, MBH 39X17870AH, MBH 39X17866 AG, MBH 39X17868 AG, MBH 39X17867 AC, 32X SEB1 D, 32X ALB12 A, MBH 32XALB7, MBH C31XB20, BCS 207/1, VHG-QC-CHECK-26-100, BAM 376a. All values were confirmed in an inter-laboratory comparison using independent analytical methods.

Analytical methods applied for characterization:

Inductively coupled plasma optical emission spectrometry (ICP-OES) Flame atomic absorption spectrometry (FAAS) Inductively coupled plasma mass spectrometry (ICP-MS) X-ray Fluorescence spectrometry with energy dispersion (EDXRF)

Participating laboratories:

	Pb	Cd	Cr	Hg
ICP-OES	1, 2, 3, 7	1, 2, 3, 7	1, 2, 3, 7	1, 2, 3, 7
FAAS	2, 4	2, 4	2, 4	
ICP-MS	2, 5	2, 5	5	2, 5
EDXRF	6	6		

- 1. Łukasiewicz Research Network Institute of Non-Ferrous Metals, Centre of Analytical Chemistry, Laboratory of Emission Spectrometry and Chromatography, Gliwice, Poland
- 2. Łukasiewicz Research Network Institute of Non-Ferrous Metals, Centre of Analytical Chemistry, Laboratory of Atomic Spectrometry, Gliwice, Poland
- Łukasiewicz Research Network Institute of Ferrous Metallurgy, Laboratory of Analytical Chemistry, Gliwice, Poland
- Łukasiewicz Research Network Cracow Institute of Technology, Department of Laboratories, Cracow, Poland
- 5. J.S. Hamilton Poland Sp. z o.o., Gdynia, Poland
- 6. Research and Development Center RCC Sp. z o. o., Głuchołazy, Poland
- 7. Intertek Test Hizmetleri A.S., Istanbul, Republic of Türkiye

Intended use:

The CRM is intended for establishing or checking the calibration of optical emission or X-ray spectrometers for analysis of samples of similar matrix composition (for micro-analysis is not verified).

Minimum sample size:

Materials designed for spark-OES spectrometry, XRF spectrometry: >1 mm spot size. For other analytical techniques minimum 0.5 g of the CRM is required.

Instructions for storage and use:

Storage the material in a dry and clean environment at room temperature.

Transport at normal conditions.

Before every use, the surface of the CRM must be prepared by milling or turning on a lathe. Chips for wet chemical analysis must be taken by turning of the sample surface. Overheating of the material during preparation should be avoided. Samples should be prepared in the same way as the CRM. Such preparation does not result in change of certified values.

Brief description of the production and certification process:

The CRM CCC2 was made by melting of all components in a crucible induction furnace and by casting into special moulds protecting elimination of segregation of the components during solidification. Cast was done by the Łukasiewicz Research Network — Cracow Institute of Technology. Homogeneity testing were carried out taking into account about 30% of the produced material. Investigations were carried out using optical emission spectrometry with low voltage spark and XRF techniques. Homogeneity was estimated statistically with application of the ANOVA.

The certification of CCC2 is valid 30 years, within the measurement uncertainties specified, provided the CRM is handled in accordance with the instructions given in this certificate.

Expired date:

30 years

Certificate Revision History: 30th of November 2022 (original certificate date)

Since 2018 our production of the certified reference materials is carried out in accordance with requirements of the ISO 17034 standard.

The Łukasiewicz Research Network —Institute of Non-Ferrous Metals holds an accreditation of the Polish Centre for Accreditation as a reference material producer according to ISO/IEC 17034 with certificate number RM 006.

Contact:

Łukasiewicz Research Network —Institute of Non-Ferrous Metals Centre of Analytical Chemistry Sowińskiego 5 44-100 Gliwice, Poland

e-mail: crm@imn.lukasiewicz.gov.pl

phone: +48 32 23 80 408

www.reference-materials.com

