

CERTIFICATE OF REFERENCE MATERIAL

ZG
Zn-Al alloy

The assigned certified values¹ and uncertainties²

	Al	Cu	Mg	Pb	Cd	Sn	Fe	Ni	
ZG1	%			mg/kg					
	3.065	1.336	0.0742	90.2	4.84	68.1	82.5	67.0	
	±0.047	±0.021	±0.0014	±8.3	±0.98	±6.8	±3.3	±4.9	
ZG2	%			mg/kg					mg/kg
	3.560	0.718	0.0485	65.3	48.8	48.3	-	25.3	
	±0.051	±0.022	±0.0016	±3.7	±1.7	±12	-	±6.8	
ZG3	%			mg/kg			%	mg/kg	
	4.003	0.109	0.0283	33.1	27.50	6.68	0.0110	10.5	
	±0.076	±0.012	±0.0027	±3.3	±0.81	±0.87	±0.0011	±1.8	
ZG4	%	mg/kg			%	mg/kg	%	mg/kg	
	4.643	-	55	12.6	0.0108	20.6	0.0163	4.2	
	±0.089	-	±15	±3.0	±0.0016	±5.5	±0.0035	±1.0	

¹ Unweighted mean value of the means of accepted sets of data, each set being obtained in a different laboratory and/or with a different method of determination.

² The certified uncertainty is the expanded uncertainty with a coverage factor k=2, corresponding to a level of confidence of about 95%.

Not certified values

	Si	Cu
ZG1	%	
	0.036	-
ZG2	%	
	0.024	-
ZG3	%	
	0.010	-
ZG4	mg/kg	
	47	89

Value for Si and Cu are presented as informative because the values were reported only by two laboratories

Signature
SIEĆ BADAWCZA ŁUKASIEWICZ-
INSTYTUT METALI NIEŻELAZNYCH
DYREKTOR

dr inż. Barbara Juszczyk, MBA

Description of the material:

The certified reference materials are available in the form of 44 mm diameter and 26 mm height discs.

Traceability:

The certified values are traceable to the SI via calibration using pure metals, certified monoelement standard solutions and certified reference materials i.e. 42X Z1 (batch J), 42X Z16 (batch A), 43X Z14 (batch E), produced by MBH Analytical LTD. All values were confirmed in an inter-laboratory comparison using independent analytical methods.

Analytical methods applied for characterization:

Al – Inductively coupled plasma optical emission spectrometry (ICP-OES), Atomic absorption spectrometry (AAS)

Cu – Inductively coupled plasma optical emission spectrometry (ICP-OES), Atomic absorption spectrometry (AAS)

Pb – Inductively coupled plasma optical emission spectrometry (ICP-OES), Atomic absorption spectrometry (AAS)

Cd – Inductively coupled plasma optical emission spectrometry (ICP-OES), Atomic absorption spectrometry (AAS)

Sn – Inductively coupled plasma optical emission spectrometry (ICP-OES), Spectrophotometric

Fe – Inductively coupled plasma optical emission spectrometry (ICP-OES), Atomic absorption spectrometry (AAS)

Ni – Inductively coupled plasma optical emission spectrometry (ICP-OES), Spectrophotometric

Si – Spectrophotometric

Participating laboratories:

1. Łukasiewicz Research Network - Institute of Non-Ferrous Metals, Analytical Chemistry Department, Emission Spectrometry Laboratory, Gliwice, Poland
2. Łukasiewicz Research Network - Institute of Non-Ferrous Metals, Analytical Chemistry Department, Atomic Spectrometry Laboratory, Gliwice, Poland
3. Walcownia Metali Nieżelaznych "Łabędy" S.A., Gliwice, Poland
4. Huta Cynku „Miasteczko Śląskie” S.A., Miasteczko Śląskie, Poland

Intended use:

The CRM is intended for establishing or checking the calibration of chemical analysis methods, for validation and to demonstrate results traceability of samples with similar matrix composition (not verified for micro-analysis).

Minimum sample size:

Minimum 0.5 g of the CRM is required.

Instructions for storage and use:

Storage the material in a dry and clean environment at room temperature.

Transport at normal conditions.

Overheating of the material during preparation should be avoided. Samples should be prepared in the same way as the CRM. Such preparation does not result in change of certified values.

Brief description of the production and certification process:

The material was produced by Łukasiewicz - IMN. Homogeneity investigations were made taking into account about 30% of the material produced. Investigations were carried out using optical emission spectrometry with low voltage spark excitation source (spark-OES). Homogeneity was estimated statistically with analysis of variance (ANOVA).

The certification of ZG is valid 50 years, within the measurement uncertainties specified, provided the CRM is handled in accordance with the instructions given in this certificate.

Expired date:

50 years

Certificate Revision History: July 2001 (original certificate date); 30th of November 2024 (additional information about: expanded uncertainties, traceability, participating laboratories, methods used for certification, minimum sample size, instruction for storage and use and expire date was added, change of graphic design)

Since 2018 our production of the certified reference materials is carried out in accordance with requirements of the ISO 17034 standard.

The Łukasiewicz Research Network —Institute of Non-Ferrous Metals holds an accreditation of the Polish Centre for Accreditation as a reference material producer according to ISO/IEC 17034 with certificate number RM 006.

Contact:

Łukasiewicz Research Network —Institute of Non-Ferrous Metals
Centre of Analytical Chemistry
Sowińskiego 5 Street
44-100 Gliwice, Poland
e-mail: crm@imn.lukasiewicz.gov.pl
phone: +48 32 23 80 408
www.reference-materials.com

